


PDS No. 6123xx	PRODUCT DATA SHEET			Page 1 of 1
Revision 05	Serum Pipette			 greiner bio-one
	Greiner Item-No. 6123xx			
Valid for Item-No.:	612301	612361 (sterile)	612362 (sterile)	

1.	Description / Specification	
1.1	Description	Serum pipette with graduation (scale 0,5 – 2,5 ml) 612301: standard 612361, -362: sterile
1.2	Dimensions	Length: 153 mm
1.3	Volume	Working volume: 2,5 ml
1.4	Material / Resin	LDPE (Low Density Polyethylene)
1.5	Colour	Translucent
1.6	Sterilization	612361, -362: SAL 10 ⁻³ 612301: no
1.7	Quality Control	<u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	-50°C to +80°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: www.gbo.com/bioscience →Products →Literature →Technical Information→Chemical Resistance of Resins
2.6	Shelf life	612361, -362: 5 years after month of production 612301: -
2.7	Other Information	-

3.	Packaging	612301	612361	612362
3.1	Pieces / Bag	500	1	25
3.2	Pieces / Box	1500	800	1000
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-	-	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 04	Date 2 March 2011	Date 3 March 2011	Date 3 March 2011	
Date 30.06.2009	Name S. Kaelberer / S. Jud	Name Dr. L. Marchetti	Name A. Schulz	